

Certificate

Traceability of Manufacturers to the IFCC Reference Method for HbA1c

This certifies that **Bayer Corporation** using **DCA 2000**, uses calibrators supplied by the IFCC Network to get traceable to the IFCC Reference Method and participates in the Monitoring Programme to demonstrate traceability. In the Monitoring Programme of 2005 the following performance was seen:

Deviation from IFCC-target	at 3 % HbA1c %	:	0.1%
	at 6 % HbA1c %	:	0.0%
	at 9 % HbA1c %	:	0.0%
Reproducibility, coefficient of variation			1.7%
Linearity, correlation coefficient			0.9981

Date of issue: 28 October 2005

Certification expires: 31 December 2006


Chair IFCC Working Group


IFCC Network Coordinator